

# Vaporiser Calibration Report

## Overview

|                        |  |         |
|------------------------|--|---------|
| Vaporiser Serial, Type | BEJK 12345 T# 00001, Tec 7 SEVOFLURANE |         |
| Customer               | ABC Hospital                           |         |
| Environment            | 21.00°C, 101353 Pa                     |         |
| Operator               | Mary Smith                             |         |
| Carrier A              | Oxygen                                 | 100.00% |
| Carrier B              | (none)                                 | 0.00%   |
| Agent                  | Sevoflurane                            | 0.00%   |

## Results

| Setting | 5.00 lpm |
|---------|----------|
| 0.00%   | 0.00     |
| 0.60%   | 0.55     |
| 1.00%   | 1.01     |
| 2.00%   | 2.01     |
| 3.00%   | 3.02     |
| 4.00%   | 4.09     |
| 5.00%   | 5.02     |
| 6.00%   | 5.95     |
| 7.00%   | 7.14     |
| 8.00%   | 8.01     |

This vaporizer has been serviced, calibrated, and certified to be within the correct operating criteria.

|                  |
|------------------|
| QUALITY CONTROL  |
| FINAL INSPECTION |
| DATE <i>N/A</i>  |
| INITIAL          |

|                 |
|-----------------|
| FINAL LEAK TEST |
| PASS            |
| DATE <i>N/A</i> |
| INITIAL         |